

# HiTrap IMAC HP 1 ml and 5 ml

HiTrap™ IMAC HP is prepacked with IMAC Sepharose™ High Performance. The column is ideal for purification of histidine-tagged proteins and other proteins/peptides by immobilized metal ion affinity chromatography (IMAC).

The design of the column, together with the high-performance Sepharose matrix, provides fast, simple and easy separations in a convenient format.

HiTrap IMAC HP columns can be operated with a syringe, peristaltic pump, or liquid chromatography systems such as ÄKTAdesign™.

## Caution

Metal ions used in IMAC may be toxic and/or cause allergic reactions.



<b>Code No.</b>	<b>Product</b>	<b>No. supplied</b>
17-0920-03	HiTrap IMAC HP 1 ml	5 × 1 ml
17-0920-05	HiTrap IMAC HP 5 ml	5 × 5 ml

### **Connector kit**

<b>Connectors supplied</b>	<b>Usage</b>	<b>No. supplied</b>
1/16" male/luer female	Connecting syringe to top of HiTrap column	1
Tubing connector flangeless/M6 female	Connecting tubing (e.g. Peristaltic Pump P1) to bottom of HiTrap column*	1
Tubing connector flangeless/M6 male	Connecting tubing (e.g. Peristaltic Pump P1) to top of HiTrap column†	1
Union 1/16" female/M6 male	Connecting original FPLC™ System through bottom of HiTrap column	1
Union M6 female/1/16" male	Connecting original FPLC System through top of HiTrap column	1
Stop plug female, 1/16"	Sealing bottom of HiTrap column	2, 5 or 7

\* Union 1/16" female/M6 male is also needed.

† Union M6 female/1/16" male is also needed.

## **Table of contents**

1. Introduction	3
2. Product description	3
3. General considerations	6
4. Operation	7
5. Optimization	12
6. Stripping and re-charging	13
7. Cleaning-in-place (CIP)	13
8. Scaling-up	14
9. Storage	14
10. Troubleshooting	15
11. Further information	18
12. Ordering information	18

# 1. Introduction

Immobilized metal ion affinity chromatography (IMAC) is a widely used separation method for purifying a broad range of proteins and peptides, either in capture steps or in subsequent intermediate or polishing purification steps. IMAC is based on the specific interaction between immobilized metal ions and certain amino acid side chains exposed on the surface of proteins (mainly His and to a lesser extent Cys and Trp). The strength of interaction with immobilized metal ions is dependent on the type, number and spatial distribution of the amino acid side chains, and on the nature of the metal ion used. The chromatographic operating conditions (pH, type and concentration of salt, additives, etc.) also contribute to the interaction observed.

The presence of several adjacent histidines, such as a (histidine)<sub>6</sub>-tag, increases the affinity for immobilized metal ions and generally makes the histidine-tagged protein the strongest binder among other proteins in an *E. coli* extract, or other samples.

## 2. Product description

HiTrap IMAC HP 1 ml and 5 ml columns are prepacked with IMAC Sepharose High Performance, which consists of 34 µm highly cross-linked agarose beads with a covalently immobilized chelating group. The small bead size gives high resolution and narrow peaks with a high concentration of the purified components.

The medium can easily be charged with Ni<sup>2+</sup>, Co<sup>2+</sup>, Zn<sup>2+</sup>, Cu<sup>2+</sup>, Fe<sup>3+</sup> or other metal ions.

HiTrap IMAC HP columns are made of biocompatible polypropylene that does not interact with biomolecules. Columns are delivered with a stopper on the inlet and a snap-off end on the outlet. The columns have porous top and bottom frits that allow high flow rates.

Columns can be operated with either a syringe and the supplied adaptor, a peristaltic pump, or a chromatography system such as ÄKTAdesign.

**Note:** HiTrap columns cannot be opened or refilled.

**Table 1.** HiTrap IMAC HP characteristics.

Matrix	Highly cross-linked spherical agarose, 6%
Average particle size	34 $\mu$ m
Dynamic binding capacity*	At least 40 mg (histidine) <sub>6</sub> -tagged protein/ml medium when charged with Ni <sup>2+</sup>
Metal ion capacity:	Approx. 15 $\mu$ mol Ni <sup>2+</sup> /ml medium
Column volumes	1 ml or 5 ml
Column dimensions	i.d. $\times$ H: 0.7 $\times$ 2.5 cm (1 ml) 1.6 $\times$ 2.5 cm (5 ml)
Recommended flow rate	1 and 5 ml/min for 1 and 5 ml column, respectively
Max. flow rates	4 and 20 ml/min for 1 and 5 ml column, respectively
Max back pressure†	0.3 MPa, 3 bar
Compatibility during use	See Table 2
Chemical stability‡	0.01 M HCl, 0.1 M NaOH. Tested for one week at 40 °C. 1 M NaOH, 70% acetic acid. Tested for 12 h. 2% SDS. Tested for 1 h. 30% 2-propanol. Tested for 30 min.
Avoid in buffers	Chelating agents, e.g. EDTA, EGTA, citrate (see Table 2)
pH stability‡	short term (< 2 h): 2–14 long term (< one week): 3–12
Storage	20% ethanol at 4 to 30 °C

- \* Conditions for determining dynamic binding capacity:  
 Samples: 1 mg/ml (histidine)<sub>6</sub>-tagged pure proteins (M<sub>r</sub> 28 000 or 43 000) in binding buffer (capacity at 10% breakthrough) or (histidine)<sub>6</sub>-tagged protein (M<sub>r</sub> 28 000) bound from *E. coli* extract  
 Column volume: 0.25 ml or 1 ml, respectively  
 Flow rate: 0.25 ml/min or 1 ml/min, respectively  
 Binding buffer: 20 mM sodium phosphate, 0.5 M NaCl, 5 mM imidazole, pH 7.4  
 Elution buffer: 20 mM sodium phosphate, 0.5 M NaCl, 0.5 M imidazole, pH 7.4

**Note:** Dynamic binding capacity is metal-ion- and protein-dependent.

† H<sub>2</sub>O at room temperature

‡ Medium without metal ion

**Table 2.** IMAC Sepharose High Performance charged with Ni<sup>2+</sup> is compatible with the following compounds, at least at the concentrations given.

Reducing agents*	5 mM DTE 5 mM DTT 20 mM β-mercaptoethanol 5 mM TCEP (Tris[2-carboxyethyl] phosphine) 10 mM reduced glutathione
Denaturing agents	8 M urea† 6 M guanidine hydrochloride‡
Detergents	2% Triton™ X-100 (nonionic) 2% Tween™ 20 (nonionic) 2% NP-40 (nonionic) 2% cholate (anionic) 1% CHAPS (zwitterionic)
Other additives	500 mM imidazole 20% ethanol 50% glycerol 100 mM Na <sub>2</sub> SO <sub>4</sub> 1.5 M NaCl 1 mM EDTA‡ 60 mM citrate‡
Buffer substances	50 mM sodium phosphate, pH 7.4 100 mM Tris-HCl, pH 7.4 100 mM Tris-acetate, pH 7.4 100 mM HEPES, pH 7.4 100 mM MOPS, pH 7.4 100 mM sodium acetate, pH 4†

\* See Notes and blank run, under "Column preparation".

† Tested for 1 week at 40 °C.

‡ The strong chelator EDTA has been used successfully in some cases at 1 mM (with histidine tagged proteins). Generally, chelating agents should be used with caution (and only in the sample, not the buffers). Any metal-ion stripping may be counteracted by addition of a small excess of MgCl<sub>2</sub> before centrifugation/filtration of the sample. Note that stripping effects may vary with the applied sample volume.

### 3. General considerations

HiTrap IMAC HP is supplied without bound metal ions and thus needs to be charged with a suitable ion before use. The choice of metal ion is dependent on the type of application and the protein to be purified. The metal ions predominantly used in IMAC are  $\text{Cu}^{2+}$ ,  $\text{Zn}^{2+}$ ,  $\text{Co}^{2+}$ , and  $\text{Ni}^{2+}$ . For choice of metal ion, see "Optimization".

We recommend protein binding at neutral to slightly alkaline pH (pH 7–8) in the presence of 0.5–1.0 M NaCl. Sodium phosphate buffers are often used. Tris-HCl can also be used, but should be avoided in cases where the metal ion-protein affinity is very low, since it may reduce binding strength. Avoid chelating agents such as EDTA or citrate in buffers, see Table 2.

Including salt e.g. 0.5–1.0 M NaCl in the buffers and samples eliminates ion-exchange effects, but can also have a marginal effect on the retention of proteins.

Imidazole at low concentrations is commonly used in the sample as well as in the binding/wash buffer to minimize binding of unwanted host cell proteins. At somewhat higher concentrations, imidazole may decrease the binding of histidine-tagged proteins. The imidazole concentration must therefore be optimized to ensure the best balance of high purity (low binding of unwanted host cell proteins) and high yield (binding all of the histidine-tagged protein), see "Optimization" for further information.

The most frequently used elution procedure for histidine-tagged proteins is based on a competitive displacement by imidazole. As alternatives to imidazole elution, target proteins can be eluted from the medium by several other methods or combinations of methods – for example, low-pH elution in the range of pH 7.5 to pH 4. Below pH 4, metal ions may be stripped off the medium.

**Note:** If the proteins are sensitive to low pH, we recommend collecting the eluted fractions in tubes containing 1 M Tris-HCl, pH 9.0 (60–200  $\mu\text{l}/\text{ml}$  fraction) to quickly restore the pH to neutral.

Chelating agents such as EGTA or EDTA can be used to elute proteins by stripping the metal ions from the medium. The target protein pool will then contain the metal ions, which can be removed by desalting on a HiTrap Desalting, a PD-10 Desalting Column, or HiPrep™ 26/10 Desalting, (see Table 3).

Elution of bound proteins with ammonium chloride or histidine has also been reported.

Leakage of  $\text{Ni}^{2+}$  and  $\text{Co}^{2+}$  from IMAC Sepharose High Performance is very low under all normal conditions. For applications where extremely low leakage during purification is critical, it can be decreased further by performing a blank run, see "Column preparation".

Likewise, a blank run should also be performed before applying buffers/samples containing reducing agents, see "Column preparation".

Whatever conditions are chosen, HiTrap IMAC HP columns can be operated with a syringe, peristaltic pump, or chromatography system.

**Note:** If Peristaltic Pump P-1 is used, the maximum flow rate that can be run on a HiTrap IMAC HP 1 ml column is 3 ml/min.

## 4. Operation

### Buffer preparation

Water and chemicals used for buffer preparation should be of high purity. High purity imidazole gives very low or no absorbance at 280 nm. Filter buffers through a 0.22  $\mu\text{m}$  or a 0.45  $\mu\text{m}$  filter before use.

If the recombinant histidine-tagged proteins are expressed as inclusion bodies, include 6 M Gua-HCl or 8 M urea in all buffers and sample. Refolding of the denatured proteins, on-column or after elution, is protein-dependent.

### Recommended buffers

Binding buffer: 20 mM sodium phosphate, 0.5 M NaCl, 20–40 mM imidazole, pH 7.4.  
The optimal imidazole concentration is protein-dependent, 20–40 mM is suitable for many histidine-tagged proteins when using  $\text{Ni}^{2+}$  or  $\text{Co}^{2+}$ .

Elution buffer: 20 mM sodium phosphate, 0.5 M NaCl, 500 mM imidazole, pH 7.4.  
The imidazole concentration required for elution is protein-dependent.

**Note:** Compared to histidine-tagged proteins, untagged, naturally occurring proteins bind immobilized metal ions with lower affinity. Thus, the concentrations of imidazole that should be used with untagged proteins are lower than the above recommended, both for binding (sometimes no imidazole needs to be added) and elution.

An alternative, especially for untagged target proteins, is elution at low pH, e.g. a linear gradient from pH 7.4 to pH 4. Like imidazole elution buffer, also low-pH elution buffers should contain 0.5 M NaCl. Example: First stepwise elution with 50 mM sodium acetate, 0.5 M NaCl, pH 6, followed by a linear gradient to 50 mM sodium acetate, 0.5 M NaCl, pH 4.

## Column preparation

1. Prepare a 0.1 M solution of the chosen metal ion in distilled water. Salts of chlorides, sulfates, etc. can be used; e.g. 0.1 M  $\text{CuSO}_4$  or 0.1 M  $\text{NiSO}_4$  are commonly used.

**Note:** Take extra precautions when working with  $\text{Fe}^{3+}$ . In neutral solutions,  $\text{Fe}^{3+}$  is easily reduced and forms compounds that can be difficult to dissolve. Immobilize  $\text{Fe}^{3+}$  at low pH, approx. pH 3, to avoid precipitation of insoluble compounds.

2. Fill the syringe or pump tubing with distilled water. Remove the stopper from the column and connect it to the syringe (with the adaptor provided), or to the pump tubing 'drop-to-drop' to avoid introducing air into the system.
3. Remove the snap-off end at the column outlet.
4. Wash the column with 5 ml or 15 ml distilled water for HiTrap IMAC 1 ml or 5 ml column, respectively. At this stage, do not use buffer to wash out the 20% solution as the metal ion can precipitate during step 5, depending on the buffer used.
5. Charge the water-washed column by loading at least 0.5 ml or 2.5 ml of the 0.1 M metal salt solution on the 1 ml and 5 ml column, respectively.



**Table 3.** Prepacked columns for desalting and buffer exchange.

Code No.	Column	Loading volume	Elution volume	Comments	Application
17-1408-01	HiTrap Desalting	0.1–1.5 ml	1.3–4.0 ml	Prepacked with Sephadex™ G-25 Superfine. Requires a syringe or pump to run.	For desalting and buffer exchange of protein extracts ( $M_r > 5000$ ).
17-5087-01	HiPrep 26/10 Desalting	Up to 15 ml	15–20 ml	Prepacked with Sephadex G-25 Fine. Requires a pump to run.	For desalting and buffer exchange of protein extracts ( $M_r > 5000$ ).
17-0851-01	PD-10 Desalting	2.5 ml	3.5 ml	Prepacked with Sephadex G-25. Requires only gravity to run.	For desalting and buffer exchange of protein extracts ( $M_r > 5000$ ).
17-0855-01	NICK™	0.1 ml	0.4 ml	Prepacked with Sephadex G-25.	For separation of proteins ( $M_r > 5000$ ) and nicktranslated DNA from radiolabelled nucleotides not shorter than 120 mers, and similar separations.
17-0853-01	NAPT™-5	0.5 ml	1.0 ml	Prepacked with Sephadex G-25	For purification of proteins ( $M_r > 5000$ ), DNA and oligo-
17-0854-01	NAP-10	1.0 ml	1.5 ml	DNA grade.	nucleotides greater than 10 bases in length.
17-0852-01	NAP-25	2.5 ml	3.5 ml	Requires only gravity to run.	

6. Wash with distilled water, 5 ml or 15 ml, for the 1 ml and 5 ml column, respectively.
7. The column is now ready for equilibration and sample application
- (8. **Optional:** In some cases, a blank run is recommended before using the column – see below.)

## Optional: Blank run

**Note:** IMAC Sepharose High Performance charged with  $\text{Ni}^{2+}$  is compatible with reducing agents, see Table 2. However, before applying buffer/sample including reducing agents we recommend removing any weakly bound metal ions by performing a blank run without reducing agents, see below. Do not leave HiTrap IMAC HP columns in buffers containing reducing agents when not in use.

**Note:** Leakage of  $\text{Ni}^{2+}$  and  $\text{Co}^{2+}$  ions from IMAC Sepharose High Performance is low under all normal conditions. For very critical applications, performing a blank run (as described below) before loading sample can further decrease leakage during purification. Such treatment is intended to remove any weakly bound metal ions that might otherwise be desorbed later, during elution of the bound protein.

### Performing a blank run:

Use binding buffer and elution buffer **without** reducing agents.

1. If the column has been stored in 20% ethanol after metal ion charging, wash it with 5 column volumes (CV) of distilled water.
2. Wash with 5 CV of the buffer that has been chosen for the protein elution, e.g. imidazole elution buffer for competitive elution, or low-pH elution buffer. Do not use EDTA/EGTA elution buffer for a blank run.
3. Equilibrate with 5–10 CV of binding buffer.

**Note:** Equilibration down to a very low concentration of imidazole, often used for binding of untagged proteins, may be slow. The equilibration can be monitored with the absorbance of imidazole, e.g. at 220 nm.

4. The column is now ready for sample application.

## Sample preparation

For optimal growth, induction, and cell lysis conditions, please refer to established protocols.

Adjust the sample to the composition and pH of the binding buffer by adding buffer, NaCl, imidazole, and additives (as required) from concentrated stock solutions, by diluting the sample with binding buffer, or by buffer exchange. Do not use strong bases or acids for pH-adjustment (precipitation risk). Shortly before applying the sample to the column, centrifuge it and/or filter it through 0.45 or 0.22  $\mu$ m filters.

**Note:** To minimize the co-adsorption of unwanted host cell proteins, it is essential to include imidazole at a low concentration in the sample and binding buffer, see "Optimization".

## Purification

---

**Please read the section "General considerations" before starting the purification**

---

1. After the column preparation, equilibrate with at least 5 column volumes (CV) of binding buffer. Recommended flow rates are 1 ml/min or 5 ml/min for the 1 and 5 ml columns, respectively.

**Note:** In some cases, we recommend a blank run before final equilibration/sample application – see "Column preparation".

2. Apply the pretreated sample using a syringe or pump.
3. Wash with binding buffer until the absorbance reaches a steady baseline (generally, at least 10–15 CV).
4. Elute the bound proteins with elution buffer, stepwise or with a linear gradient. Five CVs are usually sufficient if the protein of interest is eluted with one step. A shallow gradient, e.g. a linear gradient over 20 CV or more, may separate proteins with similar binding strengths.

**Note:** If imidazole needs to be removed from the eluted protein, use HiTrap Desalting, a PD-10 Desalting Column, or HiPrep 26/10 Desalting depending on the sample volume, see Table 3.

## 5. Optimization

### Concentration of imidazole in binding/wash buffer

Imidazole at low concentrations is commonly used in binding/wash buffers to minimize the binding of unwanted host cell proteins. For the same reason, it is important to include imidazole in the sample (generally at the same concentration as in the binding/wash buffer). At somewhat higher concentrations, imidazole may decrease the binding of histidine-tagged proteins, leading to a lower yield. The concentration of imidazole must therefore be optimized to ensure the best balance of high purity (low binding of unwanted host cell proteins), and high yield (binding all of the histidine-tagged protein). This optimal concentration is different for different histidine-tagged proteins/target proteins, and is usually slightly higher for IMAC Sepharose High Performance than for similar IMAC media on the market.

Finding the optimal imidazole concentration for a specific target protein is a trial-and-error effort, but 20–40 mM in the sample as well as in the binding/wash buffer is a good starting point for many histidine-tagged proteins when using  $\text{Ni}^{2+}$  or  $\text{Co}^{2+}$ .

For untagged target proteins, the concentration of imidazole that should be used are generally much lower than for histidine-tagged proteins, both for binding (sometimes no imidazole is needed) and elution.

When maximum binding and yield of the histidine-tagged protein is the main objective, rather than purity, choose a low imidazole concentration for binding and wash, even if this concentration may lead to suboptimal purity.

Use a high purity imidazole, which has essentially no absorbance at 280 nm.

### Choice of metal ion

$\text{Ni}^{2+}$  is usually the first choice for purifying most histidine-tagged recombinant proteins and is also the metal ion most generally used. Nevertheless, it is not always possible to predict which metal ion will be the most suitable for a given protein. The strength of binding between a protein and an immobilized metal ion is dependent on several factors, including the length, position, and exposure of the affinity tag on the protein, the type of

metal ion used, and the pH, etc. of buffers. Some proteins may therefore be easier to purify with metal ions other than  $\text{Ni}^{2+}$ , e.g.  $\text{Zn}^{2+}$ ,  $\text{Co}^{2+}$ , or  $\text{Cu}^{2+}$ .

For purification of untagged proteins,  $\text{Cu}^{2+}$  ions have frequently been used. When the binding characteristics of an untagged target protein are not known, it is advisable to test also other metal ions (e.g.  $\text{Zn}^{2+}$ ,  $\text{Ni}^{2+}$ ,  $\text{Co}^{2+}$ ) in order to establish the most suitable metal ion to use. In some instances, a weak binding to a metal ion can be exploited to achieve selective elution (higher purity) of a target protein. In some special applications,  $\text{Fe}^{3+}$  and  $\text{Ca}^{2+}$  have also been used.

## 6. Stripping and re-charging

**Note:** The column does not have to be stripped and re-charged between each purification cycle if the same protein is to be purified; it may be sufficient to strip and re-charge it after 5–7 purifications, depending on the metal ion, sample properties, sample volume, target protein, etc.

**Recommended stripping buffer:** 20 mM sodium phosphate, 0.5 M NaCl, 50 mM EDTA, pH 7.4.

Strip the column by washing with at least 5–10 CVs of stripping buffer. Wash with at least 5–10 CV of binding buffer and 5–10 CV of distilled water before re-charging the column.

Re-charge the water-washed column according to the method previously described, see "Column preparation".

## 7. Cleaning-in-place (CIP)

Clean the column when an increase in back-pressure is seen or when cross-contamination between samples is to be prevented. Before cleaning, strip off the immobilized metal ions using the recommended procedure described above. Stripping, without any additional CIP procedures, may sometimes give a satisfactory cleaning effect.

After cleaning, store in 20% ethanol (wash with 5 CV) or re-charge with the preferred metal ion.

The stripped column can be cleaned by the following procedures. For difficult cases, use reversed flow direction:

- Remove ionically bound proteins by washing with several column volumes (CV) of 1.5 M NaCl. Then wash with at least 3 CV of distilled water.
- Remove precipitated proteins, hydrophobically-bound proteins, and lipoproteins by washing with 1 M NaOH, contact time usually 1–2 h (longer time may be required to inactivate endotoxins). Then wash with 3–10 CV of binding buffer, followed by 5–10 CV of distilled water.
- Remove hydrophobically bound proteins, lipoproteins, and lipids by washing the column with 5–10 CV 30% isopropanol for at least 15–20 min. Then wash with approx. 10 CV of distilled water. Alternatively, wash with 2 CV of detergent in a basic or acidic solution. Use, for example, 0.1–0.5% nonionic detergent in 0.1 M acetic acid, contact time 1–2 h. After treatment, always remove residual detergent by washing with at least 5 CV of 70% ethanol. Then wash with 3–10 CV of distilled water.

## 8. Scaling-up

Two or three HiTrap IMAC HP 1 ml or 5 ml columns can be connected in series for quick scale-up (note that back-pressure will increase).

IMAC Sepharose High Performance, the medium that is packed in HiTrap IMAC HP columns, is supplied preswollen in 25 and 100 ml lab packs (see the ordering information). An alternative scale-up strategy is thus to pack the medium in empty columns – Tricorn™ and XK columns are suitable for this purpose.

## 9. Storage

Store at 4 to 30 °C with 20% ethanol in the columns.

## 10. Troubleshooting

The following tips may be of assistance. If you have further questions about HiTrap IMAC HP columns, please visit [www.gelifsciences.com/hitrap](http://www.gelifsciences.com/hitrap) or contact technical support or your local representative.

**Note:** Proteins generally unfold when using high concentrations of urea or Gua-HCl. Refolding on-column (or after elution) is protein-dependent.

To minimize sample dilution, add solid urea or Gua-HCl.

Samples containing urea can be analyzed directly by SDS-PAGE whereas samples containing Gua-HCl must be buffer-exchanged to a buffer with urea before SDS-PAGE.

### Column has clogged:

- Cell debris in the sample may clog the column. Clean the column according to Cleaning-in-place procedures.
- Centrifuge and/or filter the sample through a 0.22 µm or 0.45 µm filter shortly before column application, see "Sample preparation".

### Sample is too viscous:

- If the lysate is very viscous due to a high concentration of host nucleic acid, continue sonication until the viscosity is reduced, and/or add DNase I to 5 µg/ml, Mg<sup>2+</sup> to 1 mM, and incubate on ice for 10–15 min. Alternatively, draw the lysate through a syringe needle several times.

### Protein is difficult to dissolve or precipitates during purification:

- The following additives may help: 2% Triton X-100, 2% Tween 20, 2% NP-40, 2% cholate, 1% CHAPS, 1.5 M NaCl, 50% glycerol, 20 mM β-mercaptoethanol, 1–3 mM DTT or DTE (up to 5 mM is possible depending on the sample and the sample volume), 5 mM TCEP, 10 mM reduced glutathione, 8 M urea, or 6 M Gua-HCl. Mix gently for 30 min to aid solubilization of the tagged protein (inclusion bodies may require longer mixing).

**Note:** Triton X-100 and NP-40 (but not Tween) have a high absorbance at 280 nm. Furthermore, detergents cannot be easily removed by buffer exchange.

## No histidine-tagged protein in the purified fractions:

- **Elution conditions are too mild (histidine-tagged protein still bound):** Elute with an increasing imidazole gradient or decrease pH to determine optimal elution conditions.
- **Protein has precipitated in the column:** For the next experiment, decrease the amount of sample, or decrease protein concentration by eluting with a linear imidazole gradient instead of stepwise elution. Try detergents, change NaCl concentration, or elute under denaturing (unfolding) conditions (add 4–8 M urea or 4–6 M Gua-HCl).
- **Nonspecific hydrophobic or other interactions:** Add a nonionic detergent to the elution buffer (e.g. 0.2% Triton X-100) or change the NaCl concentration.
- **Concentration of imidazole in the sample and/or binding buffer is too high:** The protein is found in the flowthrough. Decrease the imidazole concentration.
- **Target protein may not be histidine-tagged as expected:** Verify the DNA sequence of the gene. Analyze samples taken before and after induction of expression with, for example, anti-histidine antibodies in Western blotting.
- **Histidine-tag may be insufficiently exposed:** The protein is found in the flowthrough; perform purification of unfolded protein in urea or Gua-HCl as for inclusion bodies.
- **Buffer/sample composition is incorrect:** The protein is found in the flowthrough material. Check pH and composition of sample and binding buffer. Ensure that chelating or strong reducing agents are not present in the sample at too a high concentration, and that the concentration of imidazole is not too high.

SDS-PAGE of samples collected during preparation of the bacterial lysate may indicate that most of the histidine-tagged protein is located in the



centrifugation pellet. Possible causes and solutions are:

- **Sonication insufficient:** Check cell disruption by microscopic examination or monitor by measuring the release of nucleic acids at 260 nm. Adding lysozyme (up to 0.1 volume of a 10 mg/ml lysozyme solution in 25 mM Tris-HCl, pH 8.0) prior to sonication may improve results. Avoid frothing and overheating as this may denature the target protein. Over-sonication can also lead to co-purification of host proteins with the target protein.
- **Protein is insoluble (inclusion bodies):** The protein can usually be solubilized (and unfolded) from inclusion bodies using common denaturants such as 4–6 M Gua-HCl, 4–8 M urea, or strong detergents. Prepare buffers containing 20 mM sodium phosphate, NaCl, 8 M urea, or 6 M Gua-HCl, and suitable imidazole concentrations, pH 7.4–7.6. Buffers with urea should also include 500 mM NaCl. Use these buffers for sample preparation, as binding buffer and as elution buffer. For sample preparation and binding buffer, use 10–20 mM imidazole or the concentration selected during optimization (including urea or Gua-HCl).

## **The protein is collected but is not pure (multiple bands on SDS polyacrylamide gel):**

- **Partial degradation of tagged protein by proteases:** Add protease inhibitors (use EDTA with caution, see Table 2).
- **Contaminants have high affinity for the metal ions used:** Optimize the concentration of imidazole in the sample and binding buffer. Elute stepwise or with a linear imidazole gradient to determine the optimal imidazole concentrations to use for binding and washing; add imidazole to the sample to the same concentration as in the binding buffer. Wash before elution with binding buffer containing as high a concentration of imidazole as possible, without causing elution of the target protein. A shallow imidazole gradient (20 column volumes or more) may separate proteins with similar binding strengths.

If optimized conditions do not remove contaminants, further purification by ion exchange chromatography (HiTrap Q HP or HiTrap SP HP) and/or gel filtration (Superdex™ Peptide, Superdex 75 or Superdex 200) may be necessary.

- **Contaminants are associated with tagged proteins:** Add detergent and/or reducing agents before sonicating cells, or shortly afterwards if foaming is a problem. Increase detergent levels (e.g. up to 2% Triton X-100 or 2% Tween 20) change the NaCl concentration, or add glycerol (up to 50%) to the wash buffer to minimize non-specific interactions.
- **Try another metal ion:** The metal ion used for purification may not be the most suitable, see "Optimization".

## Histidine-tagged protein is eluted during sample loading/wash:

- **Buffer/sample composition is incorrect:** Check pH and composition of sample and binding buffer. Ensure that chelating or strong reducing agents are not present in the sample at too high a concentration, and that the concentration of imidazole is not too high.
- **Histidine-tag is partially obstructed:** Purify under denaturing conditions (use 4–8 M urea or 4–6 M Guan-HCl).
- **Column capacity is exceeded:** Join two or three HiTrap IMAC HP 1 ml columns together or change to a HiTrap IMAC HP 5 ml column.

## 11. Further information

Visit [www.gelifesciences.com/hitrap](http://www.gelifesciences.com/hitrap) for further information.

Several handbooks also contain useful information, see Ordering information.

## 12. Ordering information

Product	No. supplied	Code No.
HiTrap IMAC HP	5 × 1 ml	17-0920-03
HiTrap IMAC HP	5 × 5 ml	17-0920-05

Related products	No. supplied	Code No.
HisTrap HP	5 × 1 ml	17-5247-01
HisTrap HP	100 × 1 ml	17-5247-05

HisTrap HP	5 × 1 ml	17-5248-01
HisTrap HP	5 × 5 ml	17-5248-02
HisTrap HP	100 × 5 ml *	17-5248-05
IMAC Sepharose High Performance	25 ml	17-0920-06
IMAC Sepharose High Performance	100 ml	17-0920-07
HiTrap Desalting	5 × 5 ml	17-1408-01
HiTrap Desalting	100 × 5 ml *	11-0003-29
PD-10 Desalting Column	30	17-0851-01
HiPrep 26/10 Desalting	1 × 53 ml	17-5087-01
HiPrep 26/10 Desalting	4 × 53 ml	17-5087-02

\* Pack size available by special order.

Accessories	No. Supplied	Code No.
1/16" male/luer female*	2	18-1112-51
Tubing connector flangeless/M6 female*	2	18-1003-68
Tubing connector flangeless/M6 male*	2	18-1017-98
Union 1/16" female/M6 male*	6	18-1112-57
Union M6 female /1/16" male*	5	18-3858-01
Union luerlock female/M6 female	2	18-1027-12
HiTrap/HiPrep, 1/16" male connector for ÄKTAdesign	8	28-4010-81
Stop plug female, 1/16" <sup>†</sup>	5	11-0004-64
Fingertight stop plug, 1/16" <sup>‡</sup>	5	11-0003-55

\* One connector included in each HiTrap package.

<sup>†</sup> Two, five, or seven stop plugs female included in HiTrap packages depending on the product.

<sup>‡</sup> One fingertight stop plug is connected to the top of each HiTrap column at delivery.

## Literature

Recombinant Protein Handbook, Protein Amplification and Simple Purification	18-1142-75
Affinity Chromatography Handbook, Principles and Methods	18-1022-29
Affinity Chromatography Columns and Media Product Profile	18-1121-86
Convenient Protein Purification, HiTrap Column Guide	18-1129-81

[www.gelifesciences.com/hitrap](http://www.gelifesciences.com/hitrap)  
[www.gelifesciences.com/protein-purification](http://www.gelifesciences.com/protein-purification)

GE Healthcare Bio-Sciences AB  
Björkgatan 30  
751 84 Uppsala  
Sweden

GE Healthcare Europe GmbH  
Munzinger Strasse 5  
D-79111 Freiburg  
Germany

GE Healthcare UK Limited  
Amersham Place  
Little Chalfont  
Buckinghamshire, HP7 9NA  
UK

GE Healthcare Bio-Sciences Corp.  
800 Centennial Avenue  
P.O. Box 1327  
Piscataway, NJ 08855-1327  
USA

GE Healthcare Bio-Sciences KK  
Sanken Bldg.  
3-25-1, Hyakunincho  
Shinjuku-ku, Tokyo 169-0073  
Japan

GE, imagination at work and GE monogram are trademarks of General Electric Company.

ÅKTA design, Drop Design, FPLC, HiPrep, HiTrap, HisTrap, NAP, NICK, Sepharose, Sephadex and Tricorn are trademarks of GE Healthcare companies.

#### Licensing information

Purification and preparation of fusion proteins and affinity peptides comprising at least two adjacent histidine residues may require a license under US patents 5,284,933 and US patents 5,310,663, including corresponding foreign patents (assigne Hoffman-La Roche, Inc).

©2006-2007 General Electric Company - All rights reserved.  
First published Jan. 2006

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.



imagination at work

28-4046-23 AB 05/2007